# LET'S GO! GUIDE

This guide will help you get started enjoying your R90 for everyday adventures. Please see your owner's manual for more in-depth instructions and installation guide.

RG

**E**/R90

# **E**|**R**90°

## COMPONENTS



## **INSTALL & SET-UP**

FOR FOLDING CHAIRS: Refer to your owner's manual for instructions on installing the removable axle tube. Depending on your camber tube size, a spacer may be included for best clamp fit.



Refer to your owner's manual for specific torque settings.

# ATTACH THE RECEIVER CLAMP

Loosely fit the receiver clamp to the axle tube. It should be snug but still able to be manipulated or rotated on the tube.



Flip the chair over to more easily access the axle tube.

### CENTER THE RECEIVER CLAMP

Use a tape measure to find the middle of the axle tube. Position the receiver clamp so there is equal distance between the clamp and each wheel.



On QUICKIE<sup>®</sup> rigid chairs, to quickly center and level the R90 receiver clamp, align the hole at the center of the axle tube with the hole at the center of the R90 receiver clamp. Threading a small Allen wrench (or similar) through the holes will easily align them.

## LEVEL THE RECEIVER CLAMP

While seated in your chair, have someone adjust the receiver clamp so it is level. Tighten all 6 screws to the correct torque setting to secure the receiver clamp.





## SET THE WHEEL UP POSITION

With the wheel up, slide the R90 into the receiver clamp. The V-lock<sup>™</sup> system will lock the R90 into place.

Slide the included wheel height gauge under the wheel. The wheel height should be no higher than the thick side of the gauge labeled "NO GO" (3/8", 9.5mm) and no lower than the thin side of the gauge labeled "GO" (1/4", 6.4mm).

### ADJUST THE WHEEL HEIGHT

Slightly loosen the two actuator fasteners (A) on the back of the R90 using a 4mm Allen wrench. Then adjust the wheel height using the manual height adjustment screw (B). Position the wheel so it rests between the "GO" and "NO GO" heights on the wheel height gauge. Retighten the two actuator bolts to the correct torque setting.



Insert the Allen wrench into the manual height adjustment first, then attach the R90 to the chair.

NOTE- If the Allen wrench interferes with the R90's ability to attach to the chair do not force a connection as you may damage the height adjustment.

CHECK– While seated in your chair, have someone use the power button to raise and lower the wheel. The wheel should make contact with the ground in the lowest position and have between 1/4 - 3/8 inch (6.4 - 9.5mm) of ground clearance when the wheel is up.







## **APP & CALIBRATION**

## DOWNLOAD THE SUNRISE INTELLIGENCE™APP

Make sure Bluetooth on your mobile device is turned on. Open the **SI** app and tap "Connect".

Select your R90. Complete the pop-up windows accepting the Sunrise Medical legal agreement.

# Enter your body weight and calibrate the R90 to your chair.

Entering your accurate weight optimizes the performance and drive parameters of the R90. You can always adjust this through the app settings.



You can download the Sunrise Intelligence<sup>™</sup> app from here:



# CONNECT THE CONTROL BOX

With a battery installed in the R90, press the STOP button to initiate the pairing process.

The R90 and Control Box will automatically search for each other. This will be indicated by alternating LED arrows on the Control Box.

When the LEDs flash 3 times, the Control Box and the R90 are connected.



# DRIVING

# LET'S DRIVE!

## Start slow. We recommend anti-tips while operating the R90.

Before you start, be sure the Control Box is connected.



Press once to lower the wheel. When you hear an audible signal from the R90 the wheel is completely lowered.



Press the button in .5-second increments to continue increasing your speed. You cannot hold the button or press it faster to speed up faster.



PRO You can manually propel the chair to your desired speed and then hit TIP the plus button once to maintain that speed.



Press to stop power to the R90.

### MANEUVERING

The R90's position directly under the chair allows vou to maneuver and turn within your chair's original footprint.

Use the R90 to accelerate to a comfortable speed. Then practice turning by applying pressure to one hand rim or the other. Practice this in both directions and at different speeds.



The R90 will sense a tight turn and decrease the speed accordingly. The indoor drive profile is more sensitive to this than the outdoor drive profile.



## HILLS AND RAMPS

Start slow. We recommend anti-tips while operating the R90.

Find a straight and moderate incline to practice.

#### **GOING UP**

Starting at least 3 meters from the base of the incline, start the R90 and gain momentum as you approach.

As you begin to ascend, the R90 will automatically increase power to the motor to help propel you up the incline. (Shifting your torso forward will improve your riding sensation and stability.)

#### **GOING DOWN**

If going down a slope, the R90 will stop power to the motor.

The R90 does not have a braking mechanism. You must perform all braking and speed control.



When you have reached the bottom of the incline, coast to your desired speed and then push the Plus Button once to maintain that speed.

## 

Using a power assist device can change your center-of-gravity or the way your chair rides. To reduce any risk of tip-over or fall, consult your healthcare provider about proper wheelchair configuration, including use of anti-tips and any other necessary adjustments.





#### **EMPOWER YOUR LIFE**



#### Scan for video instructions

Scan to view this document in Spanish, French, German, Italian, or Dutch.

Escanea para ver este documento en ESPAÑOL.

Scannez pour afficher ce document en FRANÇAIS.

Scannen, um dieses Dokument auf DEUTSCH angezeigt zu bekommen.

Inquadra qui per vedere il documento in ITALIANO.

Scan om dit document in het NEDERLANDS te bekijken.

SUNRISE MEDICAL CARES ABOUT YOUR SAFETY. Please read your owner's manual for safe operating instructions and maintenance guidelines.

800.333.4000 www.SunriseMedical.com/Empulse



Sunrise Medical (US) LLC, Fresno, CA 93727 MK-130073 Rev.C ©2.2025



IAY 13